P416

SCREENING AND DIAGNOSTIC **NEWBORN HEARING SCREENING**



Introduction

After the audit in **2020** only of the institutions had some The audit in 2023 showed the wards, 37% of community cli 20 the 3 Necessity of renovat 1 by 202 equipment – 61% of e 2. NHS coverage (%) ²etersburg (2009-2023) ¹⁰⁰ ¹ vealed than 7 years, 95,2 94.6 Untimely calibration Discontinuity in the Figure 1. Problems read The key NHS results in St Petersburg in maternity wards, improved after the audit in 2020. The NHS clinics), coverage rate increased from 89.6% (2019) Absence of certificat to 96.5% (2020) and 99.3% (2023) – fig. 2. After audit **Before audit** community clinics, Follow-up rate (to the 2nd stage of NHS) Violations in method Figure in St P€ increased on 18%. conditions, Problems with result

The Russian national model of newborn hearing screening (NHS) includes two steps: (1) OAE based screening of all the newborns in maternity wards or community clinics and (2) ABR in audiology centers for babies who failed OAE or have risk factors. NHS was implemented in 2008, but still the overall coverage of the newborns by screening is 88.6% (2022), only 19% of the children with hearing loss were diagnosed before 3 months, only 5.4% of children started their intervention program before 6 months. **Goal:** To reveal the weaknesses, problems and develop the strategy of improvement the NHS in St Petersburg. To achieve this goal the NHS system's audit program was developed and performed.



Study design

The audit was performed in 2020 and 2023. The methodology included visits in maternity wards, hospitals, community clinics to check the key indicators of NHS's good practice and collecting filled forms. During the audit workshops were also performed. Some methodological and technical problems with equipment and screening procedure were solved during the inspection. In 2020 year 78 institutions were visited by audiologists: 16 maternity wards and hospitals, 62 community clinics. In 2023 the audit was performed again, 81 institutions were inspected (21 maternity wards and hospitals, 60 community clinics).

Materials and methods The checklist included following points: - At least 1 OAE device working appropriately, - No more than 1 year since the last calibration, - Staff is certified,

Study of the state and ways of improvement the newborn hearing screening system in megalopolis

Gaziz Tufatulin^{1,2,3}, Inna Koroleva^{1,2}, Sergey Artyushkin³, Vladimir Dvoryanchikov² 1 Center Of Pediatric Audiology; 2 St Petersburg Scientific Research Institute of Ear, Nose, Throat And Speech; 3 North-Western State Medical University named after I.I.Mechnikov - St Petersburg (Russia)

> Repeated audit showed that problems with violations in methodology and testing conditions, discontinuity in the testing process and with results' documenting became less common comparing with 2020. Assessment of the screening is an important tool of improving the audiology care services. It helps to reveal problems, develop the management strategies. The audit of NHS system in St Petersburg led to increase the coverage rate and follow-up rate. Repeated audit showed the improvement in following good practice (the number of institutions who met criteria is doubled in 3-years period). Therefore, the audit should be performed on the regular basis.

- Appropriate testing conditions, - Appropriate personnel's skills, - Appropriate documenting.

1. Tufatulin GS, Koroleva IV, Artyushkin SA, Panyutina YV, Alekseenko SI. Results of the audit of the first stage newborn hearing screening in St. Petersburg. Vestn Otorinolaringol. 2023;88(2):26-30.

2. Tufatulin GS, Koroleva IV, Mefodovskaya EK. Epidemiological study of hearing impairments in children: prevalence, structure, amplification, and social factors. Vestn Otorinolaringol. 2021;86(3):28-35.



Results			
14% of inspected institutions met the checklist criteria. The rest 86% problems (fig. 1). at 36% of institutions met the good practice criteria: 33% of maternity nics (fig. 1).			
20		2023	
ion the screening devices are older		 Necessity of renovation the screening equipment – <u>53%</u> of devices are older than 7 years, 	
– 20% of devices,		 Untimely calibration, 	
testing process (6% 69% in community		 Absence of certificates in staff, 	
		 Problems with: 	
tes in staff – 27% in		 violations in methodology and testing conditions, 	
dology and testing		 discontinuity in the testing process results' documenting 	
ts' documenting.		became less common comparing with 2020.	

Conclusion

References

